


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 11 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130187
Name <u>Dan Pasquariello</u>		ID#	2620
		Date	5-05-09
A	Agency <u>Dillingham PD</u>		Phone # <u>842-5354</u>
Instrument Location <u>Dillingham PD</u>			
Alco S/N		78855	Target Value
			.085
		High Pressure	500 psi
B	Alco Test Value Average		.082
		1 st Alco	.082
		2 nd Alco	
Signature			
		CIB 5/13/09	
(OVER)			


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

MAY 05, 2009

OPERATOR'S NAME:

PASQUARIELLO, DAN

OPERATOR'S NUMBER: 2620

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VERIFICATION OF CAL

O.L. #: 1234567890

DEPT/AGENCY: DL&I

CASE/REPORT: 1234567890

TEST TYPE: V

ALCO TARGET VALUE: .085

ALCO S/N: 78855

— BREATH ANALYSIS —

.085 ADJUSTED FOR 29.83 in		17:12
ALCO TARGET	.084	17:12
BLANK TEST	.000	17:14
INTERNAL STANDARD	VERIFIED	17:14
ALCO TV 29.83 in	.082	17:14
BLANK TEST	.000	17:15
SUBJECT SAMPLE	.000	17:15
BLANK TEST	.000	17:16
ALCO TV 29.83 in	.082	17:16
BLANK TEST	.000	17:17

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

MAY 05, 2009

TIME 17:09

— DIAGNOSTIC CHECK —

COMPUTER:	OKW
PROGRAM:	OKW
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	41c
BREATH TUBE:	41c
BAROMETER:	29.85 in
FLOW DETECTOR:	OKW
PUMP	
HIGH SPEED:	OKW
DETECTOR:	OKW
FILTERS:	OKW
QUARTZ STANDARD:	OKW
CALIBRATION:	OKW

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢